

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2283

Name: RITTER 1  
Sterilizer ID: V2178198  
09/02/2022 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 166.3 F 1.1  
4:00 182.7 F 4.8  
6:00 207.1 F 10.0  
8:00 219.4 F 5.9  
10:00 228.8 F 9.9  
12:00 234.0 F 8.4  
14:00 242.0 F 12.2  
16:00 239.8 F 10.3  
18:00 252.8 F 16.8  
20:00 264.1 F 23.5  
22:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 272.1 F 28.5  
3:00 271.6 F 28.6  
4:00 272.6 F 29.1

Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 218.5 F 0.0  
10:00 216.0 F 0.0  
Dry Cycle Aborted

FILLING: 1:01  
HEATING: 22:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 13:59  
TOTAL CYCLE: 00:44:14

09/02/2022 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_