
Midmark M11 - v1.0.5
Total Cycles: 62

Name: RITTER2
Sterilizer ID: V2313228
06/01/2021 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 154.8 F 2.2
4:00 164.3 F 5.4
6:00 187.4 F 9.9
8:00 206.5 F 5.8
10:00 223.0 F 8.7
12:00 234.9 F 12.2
14:00 238.2 F 11.2
16:00 238.1 F 10.0
18:00 251.6 F 16.6
20:00 263.4 F 23.4
22:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.3
2:00 271.8 F 29.4
3:00 272.1 F 29.2
4:00 272.0 F 29.6

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 226.2 F 0.0
10:00 220.6 F 0.0
15:00 216.1 F 0.0
20:00 218.7 F 0.0
25:00 202.1 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

06/01/2021 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____