
Midmark M11 - v1.0.5
Total Cycles: 4057

Name: RITTER2
Sterilizer ID: V2313228
10/25/2024 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.2 F 0.0
2:00 118.2 F 0.0
4:00 142.4 F 0.7
6:00 171.9 F 4.0
8:00 194.8 F 7.8
10:00 215.3 F 12.2
12:00 227.5 F 6.5
14:00 231.0 F 11.1
16:00 231.6 F 7.8
18:00 244.1 F 13.0
20:00 244.0 F 12.5
22:00 255.8 F 18.7
24:00 266.7 F 25.3
26:00 273.2 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.5 F 29.9
2:00 272.0 F 29.3
3:00 271.9 F 29.7
4:00 273.0 F 29.7

Min 271.2 F 29.1
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 244.4 F 0.0
10:00 187.8 F 0.0
15:00 183.2 F 0.0
20:00 180.1 F 0.0
25:00 178.5 F 0.0
30:00 178.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:16
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:52
CYCLE COMPLETE

10/25/2024 03:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____