
Midmark M11 - v1.0.0
Total Cycles: 2940

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 181.4 F 4.6
4:00 205.3 F 9.5
6:00 216.6 F 6.4
8:00 228.7 F 9.9
10:00 231.9 F 7.7
12:00 243.8 F 13.3
14:00 243.5 F 12.1
16:00 257.0 F 19.1
18:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.3 F 28.3
3:00 272.4 F 29.2
4:00 271.8 F 28.6

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 224.5 F 0.0
10:00 244.1 F 0.0
15:00 238.3 F 0.0
20:00 227.8 F 0.0
25:00 226.7 F 0.0
30:00 229.5 F 0.0

Drying Complete

FILLING: 1:44
HEATING: 19:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

02/23/2023 03:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____