
Midmark M11 - v1.0.0
Total Cycles: 5149

Name: RITTER 1
Sterilizer ID: V2178198
12/03/2024 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 71.6 F 0.0
4:00 88.3 F 0.0
6:00 121.6 F 0.0
8:00 149.4 F 0.2
10:00 173.1 F 0.3
12:00 191.2 F 0.5
14:00 204.1 F 1.0
16:00 218.2 F 4.0
18:00 224.1 F 6.9
20:00 232.4 F 10.2
22:00 239.9 F 10.6
24:00 239.8 F 10.6
26:00 236.4 F 8.2
28:00 243.3 F 11.2
30:00 245.9 F 12.8
32:00 256.1 F 18.3
34:00 265.9 F 24.2
36:00 271.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.8 F 28.1
2:00 272.1 F 28.4
3:00 272.5 F 28.6
4:00 272.7 F 28.7

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 191.2 F 0.0
10:00 192.4 F 0.0
15:00 194.7 F 0.0
20:00 214.2 F 0.0
25:00 228.3 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 36:04
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:14:29
CYCLE COMPLETE

12/03/2024 11:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____