

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001216  
15:36 17/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:30  
STERILIZING 4:39  
136.0 C 220kPa 4:39  
136.0 C 221kPa 5:09  
136.3 C 224kPa 5:39  
136.4 C 225kPa 6:09  
136.4 C 225kPa 6:39  
136.0 C 221kPa 7:09  
136.4 C 225kPa 7:39  
136.4 C 225kPa 8:09  
136.4 C 225kPa 8:09  
VENTING 8:09  
AIR DRYING 9:03  
CYCLE COMPLETE 10:18  
-----

Digital Signature #  
0A8B5AF52AF35F88  
-----