
Midmark M11 - v1.0.0
Total Cycles: 1713

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 188.8 F 3.5
4:00 198.0 F 8.4
6:00 217.2 F 10.9
8:00 229.8 F 9.1
10:00 240.1 F 10.0
12:00 242.1 F 12.0
14:00 240.2 F 10.4
16:00 253.5 F 17.0
18:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.7 F 28.4
3:00 271.9 F 28.8
4:00 272.5 F 29.0

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 224.9 F 0.0
10:00 232.8 F 0.0
15:00 227.3 F 0.0
20:00 225.3 F 0.0
25:00 226.0 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

03/14/2022 12:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____