
Midmark M11 - v1.0.5
Total Cycles: 1228

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 197.7 F 4.4
4:00 213.4 F 9.8
6:00 225.7 F 7.3
8:00 234.0 F 11.7
10:00 235.7 F 9.8
12:00 237.9 F 8.9
14:00 250.3 F 16.1
16:00 262.7 F 23.4
18:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 271.9 F 29.9
2:00 272.4 F 29.7
3:00 271.2 F 29.1
4:00 271.9 F 30.0

Min 271.2 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 222.7 F 0.0
10:00 230.0 F 0.0
15:00 228.5 F 0.0
20:00 227.5 F 0.0
25:00 225.6 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:05
CYCLE COMPLETE

06/27/2022 12:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____