
Midmark M11 - v1.0.0
Total Cycles: 4173

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	144.1 F	0.0
2:00	142.8 F	1.4
4:00	165.4 F	4.1
6:00	177.8 F	7.6
8:00	198.1 F	11.7
10:00	211.5 F	4.1
12:00	222.6 F	8.1
14:00	233.5 F	12.4
16:00	234.0 F	8.8
18:00	241.9 F	11.6
20:00	243.7 F	12.2
22:00	254.9 F	18.0
24:00	265.0 F	24.1

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.4
1:00	272.4 F	28.9
2:00	271.4 F	28.6
3:00	272.3 F	29.0
4:00	271.4 F	28.4

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	206.8 F	0.0
	5:00	224.3 F	0.0
	10:00	227.8 F	0.0
	15:00	227.0 F	0.0
	20:00	228.8 F	0.0
Drying Complete	25:00	227.1 F	0.0
	30:00	229.9 F	0.0

FILLING: 1:06
HEATING: 25:57
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:45
CYCLE COMPLETE

01/03/2024 11:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____