
Midmark M11 - v1.0.5
Total Cycles: 1851

Name: RITTER2
Sterilizer ID: V2313228
12/15/2022 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.7 F 0.0
2:00 201.9 F 4.0
4:00 218.0 F 10.3
6:00 227.3 F 6.5
8:00 237.9 F 12.8
10:00 239.2 F 11.4
12:00 240.2 F 10.9
14:00 254.6 F 18.4
16:00 267.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.4 F 29.4
2:00 272.8 F 30.0
3:00 271.9 F 29.1
4:00 272.9 F 29.7

Min 271.8 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 205.6 F 0.0
10:00 217.0 F 0.0
15:00 219.7 F 0.0
20:00 221.4 F 0.0
25:00 194.6 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 17:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:21
CYCLE COMPLETE

12/15/2022 04:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____