
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006399
12:29 02/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:21
136.4 C 223kPa 4:21
136.1 C 222kPa 4:51
136.2 C 223kPa 5:21
136.3 C 224kPa 5:51
136.4 C 220kPa 6:21
136.4 C 225kPa 6:51
136.2 C 222kPa 7:21
135.9 C 220kPa 7:51
135.9 C 220kPa 7:51
VENTING 7:51
AIR DRYING 8:42
CYCLE COMPLETE 38:42

Digital Signature #
2EA2AAA39560D383
