
Midmark M11 - v1.0.0
Total Cycles: 4696

Name: RITTER 1
Sterilizer ID: V2178198
07/03/2024 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.9 F 0.0
2:00 129.8 F 1.7
4:00 141.7 F 4.5
6:00 165.6 F 7.8
8:00 189.2 F 11.6
10:00 218.4 F 4.9
12:00 226.0 F 9.1
14:00 241.4 F 13.4
16:00 232.2 F 9.4
18:00 237.8 F 10.0
20:00 242.7 F 12.4
22:00 254.1 F 18.2
24:00 263.8 F 24.4
26:00 272.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 271.9 F 29.4
2:00 273.0 F 29.9
3:00 273.3 F 29.8
4:00 272.3 F 29.7

Min 271.0 F 28.9
Max 274.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 230.0 F 0.0
10:00 225.4 F 0.0
15:00 224.8 F 0.0
20:00 218.4 F 0.0
25:00 230.1 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:17
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

07/03/2024 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____