

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008144  
10:26 04/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:29  
PRESSURIZING 1:33  
STERILIZING 2:33  
136.4 C 221kPa 2:33  
136.4 C 220kPa 3:03  
136.3 C 223kPa 3:33  
136.4 C 223kPa 4:03  
136.4 C 224kPa 4:33  
136.3 C 222kPa 5:03  
136.3 C 223kPa 5:33  
136.3 C 222kPa 6:03  
136.3 C 222kPa 6:03  
VENTING 6:03  
AIR DRYING 6:54  
CYCLE COMPLETE 23:55

-----  
Digital Signature #  
2F5136A0CBA106AF  
-----