
Midmark M11 - v1.0.5
Total Cycles: 682

Name: RITTER2
Sterilizer ID: V2313228
01/11/2022 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 150.6 F 2.4
4:00 167.5 F 5.9
6:00 194.9 F 10.8
8:00 214.2 F 4.9
10:00 228.7 F 10.7
12:00 230.3 F 7.9
14:00 240.0 F 11.2
16:00 245.4 F 13.7
18:00 258.4 F 20.7
20:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 30.0
2:00 271.5 F 29.2
3:00 272.4 F 30.1
4:00 272.2 F 29.6

Min 271.3 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 202.6 F 0.0
10:00 216.7 F 0.0
15:00 221.7 F 0.0
20:00 232.3 F 0.0
25:00 226.2 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

01/11/2022 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____