
Midmark M11 - v1.0.0
Total Cycles: 1388

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2021 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 73.1 F 0.0
4:00 90.2 F 0.0
6:00 133.4 F 0.0
8:00 152.6 F 0.8
10:00 177.6 F 3.5
12:00 194.6 F 6.7
14:00 209.5 F 10.5
16:00 221.0 F 7.3
18:00 227.2 F 8.0
20:00 238.0 F 12.3
22:00 235.7 F 9.1
24:00 239.8 F 9.9
26:00 246.5 F 13.4
28:00 258.5 F 19.9
30:00 269.3 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.5 F 29.1
2:00 271.7 F 28.2
3:00 272.7 F 28.9
4:00 271.7 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.7 F 0.0
5:00 213.6 F 0.0
10:00 215.2 F 0.0
15:00 215.3 F 0.0
20:00 215.0 F 0.0
25:00 220.8 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:54
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

11/23/2021 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____