
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002426
9:38 17/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:11
STERILIZING 3:14
136.0 C 220kPa 3:14
136.4 C 225kPa 3:44
136.4 C 225kPa 4:14
136.4 C 225kPa 4:44
136.4 C 225kPa 5:14
136.4 C 222kPa 5:44
136.4 C 224kPa 6:14
136.2 C 224kPa 6:44
136.2 C 224kPa 6:44
VENTING 6:44
AIR DRYING 7:31
CYCLE COMPLETE 21:41

Digital Signature #
98ACA37E1C68BCCB
