

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3635

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2023 12:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
2:00 194.7 F 4.3  
4:00 208.1 F 9.6  
6:00 221.3 F 7.1  
8:00 233.3 F 10.6  
10:00 232.2 F 7.5  
12:00 241.6 F 11.2  
14:00 246.1 F 13.4  
16:00 259.0 F 20.3  
18:00 270.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.6 F 28.8  
2:00 271.6 F 28.8  
3:00 272.4 F 28.8  
4:00 271.4 F 28.7  
  
Min 271.2 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 225.1 F 0.0  
10:00 214.7 F 0.0  
15:00 209.9 F 0.0  
20:00 203.8 F 0.0  
25:00 198.2 F 0.0  
30:00 192.5 F 0.0  
Drying Complete

FILLING: 2:15  
HEATING: 18:48  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

08/10/2023 01:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_