
Midmark M11 - v1.0.5
Total Cycles: 2881

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.8 F 0.0
2:00 164.3 F 1.7
4:00 172.2 F 5.2
6:00 194.0 F 9.5
8:00 211.0 F 8.5
10:00 221.2 F 8.5
12:00 234.2 F 13.5
14:00 237.0 F 10.7
16:00 235.2 F 8.8
18:00 248.6 F 15.2
20:00 259.8 F 21.4
22:00 270.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 30.0
2:00 272.3 F 29.4
3:00 271.4 F 29.1
4:00 272.2 F 29.9

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 215.4 F 0.0
10:00 218.0 F 0.0
15:00 218.2 F 0.0
20:00 213.7 F 0.0
25:00 215.3 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 22:60
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

09/14/2023 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____