
Midmark M11 - v1.0.0
Total Cycles: 3637

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2023 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 189.0 F 4.6
4:00 208.0 F 10.0
6:00 226.2 F 6.3
8:00 234.9 F 11.4
10:00 234.8 F 8.9
12:00 238.7 F 9.6
14:00 248.1 F 14.3
16:00 260.4 F 21.3
18:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.3 F 28.7
2:00 272.3 F 29.2
3:00 271.6 F 28.4
4:00 272.6 F 29.1

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 227.6 F 0.0
10:00 226.8 F 0.0
15:00 224.8 F 0.0
20:00 225.4 F 0.0
25:00 218.7 F 0.0
30:00 224.0 F 0.0

Drying Complete

FILLING: 2:19
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

08/10/2023 03:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____