
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002316
15:23 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:56
STERILIZING 3:03
136.4 C 225kPa 3:03
135.9 C 222kPa 3:33
136.2 C 224kPa 4:03
136.4 C 226kPa 4:33
136.1 C 221kPa 5:03
136.4 C 226kPa 5:33
136.1 C 222kPa 6:03
136.3 C 225kPa 6:33
136.3 C 225kPa 6:33
VENTING 6:33
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
6B9D6E8B46C7A271
