
Midmark M11 - v1.0.0
Total Cycles: 3137

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 171.9 F 0.0
4:00 192.6 F 0.0
6:00 212.1 F 4.9
8:00 228.5 F 9.9
10:00 232.0 F 8.3
12:00 241.7 F 12.3
14:00 240.4 F 10.7
16:00 238.8 F 9.7
18:00 253.8 F 17.0
20:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.6 F 28.6
2:00 271.8 F 28.2
3:00 272.8 F 28.6
4:00 271.7 F 27.9

Min 271.6 F 27.9
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 220.4 F 0.0
10:00 238.5 F 0.0
15:00 230.3 F 0.0
20:00 221.5 F 0.0
25:00 218.4 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

04/20/2023 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____