

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2161

Name: RITTER 1  
Sterilizer ID: V2178198  
07/21/2022 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.4 F 0.0  
2:00 160.4 F 2.5  
4:00 181.5 F 6.5  
6:00 202.7 F 11.2  
8:00 221.3 F 5.8  
10:00 234.1 F 11.1  
12:00 231.1 F 7.6  
14:00 245.0 F 13.4  
16:00 242.7 F 11.8  
18:00 254.8 F 18.1  
20:00 266.1 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.1 F 29.2  
2:00 272.7 F 29.2  
3:00 272.1 F 28.8  
4:00 271.9 F 28.9

Min 270.9 F 28.3  
Max 273.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
5:00 234.1 F 0.0  
10:00 215.3 F 0.0  
15:00 226.2 F 0.0  
20:00 232.4 F 0.0  
25:00 228.9 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:47  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:41  
CYCLE COMPLETE

07/21/2022 10:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_