
Midmark M11 - v1.0.0
Total Cycles: 4970

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.5 F 0.0
2:00 165.0 F 0.6
4:00 171.3 F 2.8
6:00 194.1 F 6.9
8:00 224.3 F 11.5
10:00 223.7 F 6.0
12:00 236.1 F 10.7
14:00 232.4 F 7.4
16:00 243.9 F 13.1
18:00 242.5 F 11.4
20:00 254.5 F 17.7
22:00 265.2 F 24.3
24:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.5 F 28.6
2:00 271.8 F 28.6
3:00 272.0 F 28.8
4:00 272.3 F 29.0

Min 271.4 F 28.6
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 221.3 F 0.0
10:00 193.2 F 0.0
15:00 207.4 F 0.0
20:00 245.9 F 0.0
25:00 231.3 F 0.0
30:00 241.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

10/07/2024 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____