
Midmark M11 - v1.0.0
Total Cycles: 3931

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 173.6 F 3.3
4:00 181.5 F 7.1
6:00 204.7 F 12.4
8:00 222.8 F 6.3
10:00 233.1 F 11.1
12:00 231.6 F 7.7
14:00 245.0 F 12.8
16:00 243.7 F 12.2
18:00 256.2 F 18.7
20:00 267.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.8
2:00 271.5 F 28.7
3:00 272.3 F 28.7
4:00 271.4 F 28.6

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 236.8 F 0.0
10:00 233.1 F 0.0
15:00 229.7 F 0.0
20:00 226.3 F 0.0
25:00 230.3 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

10/20/2023 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____