
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001101
12:49 01/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:03
PRESSURIZING 4:03
STERILIZING 5:16
136.0 C 220kPa 5:16
136.0 C 222kPa 5:46
136.2 C 220kPa 6:16
136.4 C 224kPa 6:46
136.3 C 223kPa 7:16
136.4 C 225kPa 7:46
136.3 C 223kPa 8:16
136.3 C 224kPa 8:46
136.3 C 224kPa 8:46
VENTING 8:46
AIR DRYING 9:45
CYCLE COMPLETE 16:19

Digital Signature #
A0ECA729E739D7E9
