

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5409

Name: RITTER 1  
Sterilizer ID: V2178198  
03/11/2025 04:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.6 F 0.0  
2:00 176.1 F 3.0  
4:00 187.4 F 7.0  
6:00 200.6 F 12.3  
8:00 217.1 F 5.6  
10:00 233.1 F 10.6  
12:00 233.1 F 7.2  
14:00 243.3 F 13.2  
16:00 242.4 F 11.8  
18:00 255.3 F 18.5  
20:00 267.0 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.7 F 29.2  
2:00 271.8 F 28.8  
3:00 272.7 F 29.2  
4:00 271.7 F 28.7  
  
Min 271.6 F 28.6  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 202.5 F 0.0  
10:00 215.4 F 0.0  
15:00 217.5 F 0.0  
20:00 218.5 F 0.0  
25:00 219.3 F 0.0  
30:00 224.2 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

03/11/2025 05:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_