
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002299
7:43 27/08/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:12
STERILIZING 4:10
136.0 C 220kPa 4:10
135.9 C 220kPa 4:40
135.9 C 220kPa 5:10
136.2 C 222kPa 5:40
136.2 C 221kPa 6:10
136.4 C 222kPa 6:40
136.3 C 222kPa 7:10
136.4 C 222kPa 7:40
136.4 C 222kPa 7:40
VENTING 7:40
AIR DRYING 8:33
CYCLE COMPLETE 8:36

Digital Signature #
2C7C3E351D68C2E3
