
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000726
9:57 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:44
136.4 C 222kPa 3:44
136.5 C 233kPa 4:14
136.3 C 232kPa 4:44
136.3 C 232kPa 5:14
136.1 C 231kPa 5:44
136.3 C 233kPa 6:14
136.1 C 230kPa 6:44
136.4 C 233kPa 7:14
136.4 C 233kPa 7:14
VENTING 7:14
AIR DRYING 8:07
CYCLE COMPLETE 8:09

Digital Signature #
53EB0C59007435B4
