
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000382
16:23 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:38
STERILIZING 3:45
136.4 C 232kPa 3:45
136.3 C 232kPa 4:15
136.4 C 228kPa 4:45
136.4 C 233kPa 5:15
136.4 C 233kPa 5:45
136.1 C 231kPa 6:15
136.3 C 233kPa 6:45
136.4 C 233kPa 7:15
136.4 C 233kPa 7:15
VENTING 7:15
AIR DRYING 8:06
CYCLE COMPLETE 8:27

Digital Signature #
CAB899C7BD61CB11
