
Midmark M11 - v1.0.0
Total Cycles: 4206

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2024 04:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.6 F 0.0
2:00 153.7 F 2.1
4:00 193.6 F 6.4
6:00 201.1 F 11.7
8:00 223.2 F 5.3
10:00 231.3 F 10.4
12:00 231.5 F 7.2
14:00 244.5 F 12.8
16:00 242.0 F 11.1
18:00 255.3 F 17.9
20:00 266.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.4
2:00 271.9 F 28.5
3:00 272.0 F 28.2
4:00 272.1 F 28.7

Min 271.2 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 245.5 F 0.0
10:00 223.5 F 0.0
15:00 231.7 F 0.0
20:00 225.1 F 0.0
25:00 227.4 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

01/15/2024 05:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____