

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001416  
12:42 14/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 2:56  
STERILIZING 4:00  
136.0 C 221kPa 4:00  
136.2 C 224kPa 4:30  
136.4 C 225kPa 5:00  
136.2 C 223kPa 5:30  
136.4 C 223kPa 6:00  
136.4 C 225kPa 6:30  
136.4 C 222kPa 7:00  
136.3 C 224kPa 7:30  
136.3 C 224kPa 7:30  
VENTING 7:30  
AIR DRYING 8:19  
CYCLE COMPLETE 23:19  
-----

Digital Signature #  
9AE9D7C11BE8149E  
-----