

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2918

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2023 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.0 F 0.0  
2:00 171.1 F 2.7  
4:00 185.6 F 6.8  
6:00 201.1 F 11.4  
8:00 217.2 F 5.9  
10:00 229.5 F 10.9  
12:00 230.9 F 7.7  
14:00 242.8 F 12.9  
16:00 243.2 F 12.4  
18:00 255.0 F 18.6  
20:00 265.6 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.7 F 29.7  
2:00 271.0 F 29.1  
3:00 273.0 F 30.2  
4:00 272.1 F 30.0

Min 270.9 F 29.0  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 202.1 F 0.0  
10:00 209.2 F 0.0  
15:00 212.5 F 0.0  
20:00 209.4 F 0.0  
25:00 211.6 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

09/25/2023 12:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_