
Midmark M11 - v1.0.0
Total Cycles: 3206

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2023 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.0 F 0.0
2:00 80.7 F 0.0
4:00 115.9 F 0.0
6:00 131.9 F 0.0
8:00 159.3 F 1.5
10:00 189.7 F 4.4
12:00 204.7 F 8.5
14:00 218.6 F 12.6
16:00 225.6 F 6.4
18:00 236.2 F 10.6
20:00 234.1 F 7.6
22:00 245.5 F 12.9
24:00 242.8 F 11.3
26:00 256.0 F 18.1
28:00 267.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 271.7 F 27.9
2:00 271.7 F 28.0
3:00 271.7 F 28.0
4:00 271.6 F 27.7

Min 271.6 F 27.7
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 227.9 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 29:14
STERILIZING: 4:00
VENTING: 2:45
DRYING: 1:45
TOTAL CYCLE: 00:38:46

05/05/2023 09:44 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____