
Midmark M11 - v1.0.5
Total Cycles: 4145

Name: RITTER2
Sterilizer ID: V2313228
11/20/2024 06:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 167.2 F 3.2
4:00 190.2 F 7.9
6:00 217.7 F 12.4
8:00 230.1 F 8.7
10:00 236.3 F 10.2
12:00 242.0 F 12.5
14:00 241.8 F 11.7
16:00 255.8 F 18.9
18:00 268.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 271.9 F 28.9
2:00 272.3 F 29.1
3:00 272.5 F 29.4
4:00 272.2 F 29.7

Min 271.4 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 226.7 F 0.0
10:00 224.1 F 0.0
15:00 228.2 F 0.0
20:00 223.7 F 0.0
25:00 229.1 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

11/20/2024 07:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____