

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1089

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2022 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
2:00 179.9 F 3.8  
4:00 184.9 F 8.1  
6:00 204.5 F 13.2  
8:00 220.6 F 7.8  
10:00 235.1 F 13.1  
12:00 237.2 F 10.7  
14:00 236.4 F 9.1  
16:00 250.4 F 15.9  
18:00 262.1 F 22.7  
20:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.7  
2:00 271.7 F 28.9  
3:00 272.4 F 29.8  
4:00 272.4 F 29.4

Min 271.4 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 213.4 F 0.0  
10:00 227.3 F 0.0  
15:00 219.8 F 0.0  
20:00 219.8 F 0.0  
25:00 220.3 F 0.0  
30:00 207.0 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:25  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:02  
CYCLE COMPLETE

05/10/2022 12:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_