
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 003708
11:14 27/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:08
STERILIZING 3:05
136.0 C 220kPa 3:05
136.2 C 220kPa 3:35
136.4 C 223kPa 4:05
136.4 C 224kPa 4:35
136.4 C 224kPa 5:05
136.3 C 223kPa 5:35
136.4 C 224kPa 6:05
136.4 C 224kPa 6:35
136.4 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:23
CYCLE COMPLETE 7:25

Digital Signature #
DC882B5084877561
