
Midmark M11 - v1.0.0
Total Cycles: 784

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2021 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 129.6 F 0.0
4:00 123.0 F 0.1
6:00 153.6 F 0.3
8:00 185.5 F 3.9
10:00 206.0 F 7.8
12:00 221.2 F 11.9
14:00 231.2 F 7.0
16:00 237.1 F 12.4
18:00 239.0 F 9.8
20:00 236.3 F 8.4
22:00 249.7 F 14.8
24:00 261.9 F 21.7
26:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.5
1:00 271.7 F 28.5
2:00 272.3 F 28.9
3:00 272.7 F 28.9
4:00 272.5 F 28.9

Min 271.2 F 28.4
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 234.3 F 0.0
10:00 215.5 F 0.0
15:00 219.3 F 0.0
20:00 217.2 F 0.0
25:00 222.2 F 0.0
Dry Cycle Aborted

FILLING: 0:52
HEATING: 26:44
STERILIZING: 4:00
VENTING: 2:60
DRYING: 29:03
TOTAL CYCLE: 01:03:39

04/14/2021 02:29 PM
ACCEPTED: X
REJECTED:
Operator: U01