
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005115
11:00 14/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:29
STERILIZING 3:25
136.4 C 220kPa 3:25
136.4 C 223kPa 3:55
136.2 C 223kPa 4:25
136.4 C 223kPa 4:55
136.4 C 222kPa 5:25
136.3 C 224kPa 5:55
136.3 C 223kPa 6:25
136.2 C 218kPa 6:55
136.2 C 218kPa 6:55
VENTING 6:55
AIR DRYING 7:46
CYCLE COMPLETE 16:36

Digital Signature #
82660C79C5231CCB
