
Midmark M11 - v1.0.5
Total Cycles: 88

Name: RITTER2
Sterilizer ID: V2313228
06/10/2021 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.0 F 0.0
2:00 166.3 F 2.7
4:00 186.9 F 6.2
6:00 204.0 F 10.5
8:00 220.5 F 5.8
10:00 227.3 F 9.2
12:00 236.5 F 11.1
14:00 240.6 F 11.7
16:00 239.9 F 10.6
18:00 252.2 F 16.7
20:00 263.1 F 23.1
22:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.3 F 29.4
2:00 271.4 F 28.7
3:00 271.7 F 29.4
4:00 272.5 F 29.6

Min 271.0 F 28.6
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.1 F 0.0
5:00 217.6 F 0.0
10:00 223.0 F 0.0
15:00 216.9 F 0.0
20:00 191.6 F 0.0
25:00 177.8 F 0.0
30:00 175.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:39
CYCLE COMPLETE

06/10/2021 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____