
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001503
11:41 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:33
136.0 C 220kPa 3:33
136.4 C 222kPa 4:03
136.2 C 224kPa 4:33
136.2 C 223kPa 5:03
136.4 C 226kPa 5:33
136.4 C 226kPa 6:03
136.4 C 225kPa 6:33
136.3 C 225kPa 7:03
136.3 C 225kPa 7:03
VENTING 7:03
AIR DRYING 7:53
CYCLE COMPLETE 8:03

Digital Signature #
E8E1AE76F01F26C2
