
Midmark M11 - v1.0.0
Total Cycles: 3513

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2023 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 197.2 F 3.0
4:00 207.5 F 8.2
6:00 228.7 F 13.6
8:00 232.2 F 9.6
10:00 236.6 F 8.8
12:00 244.3 F 13.3
14:00 243.5 F 12.2
16:00 256.5 F 18.9
18:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.9 F 29.1
2:00 272.8 F 29.1
3:00 272.7 F 29.3
4:00 272.0 F 29.2

Min 271.0 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 209.3 F 0.0
10:00 236.2 F 0.0
15:00 230.4 F 0.0
20:00 226.6 F 0.0
25:00 225.7 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

07/11/2023 01:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____