

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1464

Name: RITTER2  
Sterilizer ID: V2313228  
09/07/2022 02:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.0 F 0.0  
2:00 173.3 F 3.5  
4:00 190.7 F 7.5  
6:00 208.3 F 12.5  
8:00 216.1 F 6.9  
10:00 231.7 F 12.1  
12:00 233.5 F 9.1  
14:00 238.1 F 10.1  
16:00 245.7 F 13.7  
18:00 257.9 F 20.3  
20:00 268.8 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.2 F 29.9  
2:00 273.0 F 30.0  
3:00 272.6 F 29.7  
4:00 272.0 F 29.4  
  
Min 271.2 F 29.0  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 211.8 F 0.0  
10:00 225.7 F 0.0  
15:00 229.7 F 0.0  
20:00 228.7 F 0.0  
25:00 227.3 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

09/07/2022 03:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_