
Midmark M11 - v1.0.0
Total Cycles: 960

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.7 F 0.0
2:00 211.2 F 4.1
4:00 220.6 F 9.8
6:00 230.5 F 6.6
8:00 238.3 F 11.6
10:00 237.3 F 9.5
12:00 239.1 F 9.1
14:00 249.8 F 15.0
16:00 261.8 F 21.8
18:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.7
2:00 271.4 F 28.2
3:00 272.4 F 28.7
4:00 271.4 F 28.1

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 230.3 F 0.0
10:00 228.0 F 0.0
15:00 231.5 F 0.0
20:00 228.4 F 0.0
25:00 230.9 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

07/05/2021 11:55 AM
ACCEPTED: X
REJECTED:
Operator: U02