
Midmark M11 - v1.0.0
Total Cycles: 4809

Name: RITTER 1
Sterilizer ID: V2178198
08/13/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.3 F 0.0
2:00 139.8 F 0.7
4:00 159.2 F 1.1
6:00 183.1 F 3.4
8:00 203.9 F 7.1
10:00 219.3 F 11.1
12:00 220.3 F 4.9
14:00 233.7 F 9.7
16:00 236.4 F 9.9
18:00 241.9 F 11.4
20:00 237.9 F 9.3
22:00 250.5 F 15.4
24:00 261.3 F 21.4
26:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.6
2:00 272.3 F 28.9
3:00 271.6 F 28.3
4:00 272.6 F 28.7

Min 271.4 F 28.3
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 202.6 F 0.0
10:00 191.5 F 0.0
15:00 194.4 F 0.0
20:00 183.3 F 0.0
25:00 181.1 F 0.0
30:00 183.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:34
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

08/13/2024 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____