
Midmark M11 - v1.0.0
Total Cycles: 2885

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 193.5 F 4.7
4:00 208.4 F 9.6
6:00 217.6 F 7.1
8:00 231.6 F 9.6
10:00 238.1 F 8.5
12:00 244.1 F 13.2
14:00 243.9 F 12.3
16:00 257.4 F 19.4
18:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.2 F 29.0
2:00 271.2 F 28.3
3:00 272.3 F 29.2
4:00 271.8 F 28.6

Min 271.2 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 217.6 F 0.0
10:00 234.5 F 0.0
15:00 230.1 F 0.0
20:00 228.8 F 0.0
25:00 229.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

02/09/2023 05:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____