

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2051

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2022 09:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
2:00 185.2 F 3.0  
4:00 191.7 F 6.8  
6:00 222.4 F 11.3  
8:00 224.3 F 5.9  
10:00 238.0 F 11.0  
12:00 232.1 F 7.5  
14:00 245.2 F 12.8  
16:00 243.6 F 12.1  
18:00 256.0 F 18.5  
20:00 267.0 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 29.1  
2:00 272.1 F 28.6  
3:00 271.2 F 28.2  
4:00 272.3 F 29.1  
  
Min 271.2 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.1 F 0.0  
5:00 232.4 F 0.0  
10:00 238.2 F 0.0  
15:00 233.0 F 0.0  
20:00 229.0 F 0.0  
25:00 227.5 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 21:34  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:10  
CYCLE COMPLETE

06/24/2022 10:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_