
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000127
13:55 11/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:41
STERILIZING 4:41
136.0 C 220kPa 4:41
136.4 C 223kPa 5:11
136.3 C 224kPa 5:41
136.0 C 218kPa 6:11
136.3 C 224kPa 6:41
136.3 C 222kPa 7:11
136.4 C 224kPa 7:41
136.4 C 224kPa 8:11
136.4 C 224kPa 8:11
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 9:14

Digital Signature #
47D4403847B473A9
