
Midmark M11 - v1.0.0
Total Cycles: 779

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.6 F 0.0
2:00 120.2 F 0.6
4:00 132.7 F 2.6
6:00 163.7 F 5.9
8:00 217.1 F 10.7
10:00 222.9 F 4.8
12:00 225.8 F 10.1
14:00 232.3 F 7.6
16:00 243.9 F 12.8
18:00 241.9 F 11.0
20:00 254.7 F 17.8
22:00 266.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.5 F 28.3
2:00 271.9 F 28.4
3:00 271.9 F 28.7
4:00 272.4 F 29.0

Min 271.1 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 245.0 F 0.0
10:00 229.9 F 0.0
15:00 228.1 F 0.0
20:00 231.1 F 0.0
25:00 230.0 F 0.0
30:00 231.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:26
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

04/08/2021 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____