
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005766
12:37 15/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:36
STERILIZING 2:38
136.4 C 224kPa 2:38
136.1 C 218kPa 3:08
135.9 C 221kPa 3:38
136.0 C 219kPa 4:08
136.2 C 222kPa 4:38
136.4 C 224kPa 5:08
136.4 C 223kPa 5:38
136.2 C 222kPa 6:08
136.2 C 222kPa 6:08
VENTING 6:08
AIR DRYING 6:57
CYCLE COMPLETE 7:04

Digital Signature #
59EBE1A8DF5C32D5
