

-----  
Midmark M11 - v1.0.0  
Total Cycles: 632

Name: MIDMARK  
Sterilizer ID: V2178198  
01/26/2021 03:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.2 F 0.0  
2:00 156.1 F 2.8  
4:00 177.8 F 7.2  
6:00 221.7 F 13.1  
8:00 222.2 F 8.9  
10:00 235.9 F 8.6  
12:00 245.0 F 12.8  
14:00 243.1 F 11.6  
16:00 257.4 F 19.0  
18:00 269.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.1 F 28.6  
2:00 271.6 F 28.8  
3:00 272.1 F 28.4  
4:00 271.8 F 29.0  
  
Min 271.0 F 28.0  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 230.9 F 0.0  
10:00 225.4 F 0.0  
15:00 227.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:15  
HEATING: 18:53  
STERILIZING: 4:00  
VENTING: 2:53  
DRYING: 16:17  
TOTAL CYCLE: 00:43:17

01/26/2021 03:45 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01